

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110723\
 Data File : VU056097.D
 Acq On : 07 Nov 2023 14:33
 Operator : MD/SY
 Sample : 05239-07RE
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AP8RE

Quant Time: Nov 08 00:43:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 08 00:38:11 2023
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		6.245	114	368551	5.000	ug/L	0.00
28) Chlorobenzene-d5		9.415	117	366688	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.807	152	158635	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.595	65	139688	6.106	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	122.200%	
7) Chloroethane-d5		1.907	69	108230	6.972	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	139.400%#	
11) 1,1-Dichloroethene-d2		2.560	65	61890	6.141	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	122.800%	
20) 2-Butanone-d5		4.618	46	264967	70.603	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	141.200%#	
24) Chloroform-d		5.055	84	259613	5.962	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	119.200%	
26) 1,2-Dichloroethane-d4		5.695	65	122795	6.404	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	128.000%	
32) Benzene-d6		5.721	84	546062	6.200	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	124.000%	
36) 1,2-Dichloropropane-d6		6.685	67	165476	6.383	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	127.600%	
41) Toluene-d8		7.894	98	483724	5.994	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	119.800%	
43) trans-1,3-Dichloroprop...		8.177	79	60189	6.074	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	121.400%	
46) 2-Hexanone-d5		8.631	63	213363	65.691	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	131.380%#	
56) 1,1,2,2-Tetrachloroeth...		10.753	84	113943	6.158	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	123.200%#	
66) 1,2-Dichlorobenzene-d4		12.190	152	163199	6.671	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	133.400%#	
Target Compounds							Qvalue
13) Acetone		2.628	43	47579	17.191	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110723\
Data File : VU056097.D
Acq On : 07 Nov 2023 14:33
Operator : MD/SY
Sample : 05239-07RE
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AP8RE

Quant Time: Nov 08 00:43:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 08 00:38:11 2023
Response via : Initial Calibration

