

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110723\
 Data File : VU056106.D
 Acq On : 07 Nov 2023 18:12
 Operator : MD/SY
 Sample : 05281-09
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AD0

Quant Time: Nov 08 00:45:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 08 00:38:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	367460	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	371767	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	159304	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	136006	5.963	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 119.200%		
7) Chloroethane-d5	1.907	69	102606	6.629	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 132.600%#		
11) 1,1-Dichloroethene-d2	2.563	65	58396	5.811	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 116.200%		
20) 2-Butanone-d5	4.618	46	265678	71.002	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 142.000%#		
24) Chloroform-d	5.055	84	254741	5.867	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 117.400%		
26) 1,2-Dichloroethane-d4	5.695	65	121425	6.352	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 127.000%		
32) Benzene-d6	5.721	84	525260	5.883	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 117.600%		
36) 1,2-Dichloropropane-d6	6.685	67	156649	5.960	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 119.200%		
41) Toluene-d8	7.894	98	455172	5.563	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 111.200%		
43) trans-1,3-Dichloroprop...	8.177	79	57751	5.748	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 115.000%		
46) 2-Hexanone-d5	8.627	63	208829	63.417	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 126.840%		
56) 1,1,2,2-Tetrachloroeth...	10.750	84	115794	6.173	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 123.400%#		
66) 1,2-Dichlorobenzene-d4	12.190	152	164010	6.676	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 133.600%#		

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110723\
Data File : VU056106.D
Acq On : 07 Nov 2023 18:12
Operator : MD/SY
Sample : 05281-09
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AD0

Quant Time: Nov 08 00:45:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 08 00:38:11 2023
Response via : Initial Calibration

