

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110723\
 Data File : VU056105.D
 Acq On : 07 Nov 2023 17:48
 Operator : MD/SY
 Sample : 05281-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AC9

Quant Time: Nov 08 00:45:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 08 00:38:11 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	370476	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	374303	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	163410	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	132236	5.750	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 115.000%		
7) Chloroethane-d5	1.907	69	96228	6.166	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 123.400%		
11) 1,1-Dichloroethene-d2	2.563	65	57276	5.653	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 113.000%		
20) 2-Butanone-d5	4.618	46	253885	67.298	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 134.600%#		
24) Chloroform-d	5.055	84	249954	5.710	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 114.200%		
26) 1,2-Dichloroethane-d4	5.695	65	118523	6.149	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 123.000%		
32) Benzene-d6	5.721	84	513721	5.714	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 114.200%		
36) 1,2-Dichloropropane-d6	6.685	67	155016	5.858	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 117.200%		
41) Toluene-d8	7.894	98	448399	5.443	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 108.800%		
43) trans-1,3-Dichloroprop...	8.177	79	56554	5.591	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 111.800%		
46) 2-Hexanone-d5	8.631	63	207527	62.595	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 125.180%		
56) 1,1,2,2-Tetrachloroeth...	10.750	84	114692	6.072	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 121.400%#		
66) 1,2-Dichlorobenzene-d4	12.190	152	152847	6.065	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 121.400%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110723\
Data File : VU056105.D
Acq On : 07 Nov 2023 17:48
Operator : MD/SY
Sample : 05281-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AC9

Quant Time: Nov 08 00:45:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 08 00:38:11 2023
Response via : Initial Calibration

