

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110818\
 Data File : VU028015.D
 Acq On : 08 Nov 2018 10:59
 Operator : MD/SY
 Sample : PB114610TB
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB114610TB

Quant Time: Nov 08 11:44:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	188566	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	369504	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	329227	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	136885	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	159429	56.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.74%	
35) Dibromofluoromethane	4.89	113	121535	52.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.34%	
50) Toluene-d8	7.57	98	474945	52.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.54%	
62) 4-Bromofluorobenzene	10.31	95	156286	47.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.16%	
Target Compounds						
16) Acetone	2.32	43	14350	8.250	ug/l	95
20) Methylene Chloride	2.71	84	6194	2.439	ug/l	94
43) Isopropyl Acetate	5.55	43	55680	6.268	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110818\
Data File : VU028015.D
Acq On : 08 Nov 2018 10:59
Operator : MD/SY
Sample : PB114610TB
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB114610TB

Quant Time: Nov 08 11:44:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

