

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110818\
 Data File : VU028019.D
 Acq On : 08 Nov 2018 12:32
 Operator : MD/SY
 Sample : PB114610ZHE#03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB114610ZHE#03

Quant Time: Nov 08 13:16:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	190768	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	372775	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	341649	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	142998	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	163944	57.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.60%	
35) Dibromofluoromethane	4.89	113	120969	51.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.94%	
50) Toluene-d8	7.57	98	487028	53.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.28%	
62) 4-Bromofluorobenzene	10.31	95	165571	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
Target Compounds						
16) Acetone	2.32	43	14474	8.225	ug/l	95
20) Methylene Chloride	2.71	84	6135	2.388	ug/l	92
43) Isopropyl Acetate	5.55	43	54985	6.135	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110818\
Data File : VU028019.D
Acq On : 08 Nov 2018 12:32
Operator : MD/SY
Sample : PB114610ZHE#03
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB114610ZHE#03

Quant Time: Nov 08 13:16:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

