

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110818\
 Data File : VU028020.D
 Acq On : 08 Nov 2018 12:55
 Operator : MD/SY
 Sample : PB114610ZHE#04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB114610ZHE#04

Quant Time: Nov 08 13:18:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	190920	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	364360	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	333665	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	140592	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	159577	56.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.44%	
35) Dibromofluoromethane	4.89	113	118075	51.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.80%	
50) Toluene-d8	7.57	98	476923	53.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.48%	
62) 4-Bromofluorobenzene	10.31	95	161353	49.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.58%	
Target Compounds						
16) Acetone	2.32	43	14549	8.261	ug/l	99
20) Methylene Chloride	2.70	84	6001	2.334	ug/l	95
43) Isopropyl Acetate	5.55	43	53035	6.054	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110818\
Data File : VU028020.D
Acq On : 08 Nov 2018 12:55
Operator : MD/SY
Sample : PB114610ZHE#04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB114610ZHE#04

Quant Time: Nov 08 13:18:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

