

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110818\
 Data File : VU028026.D
 Acq On : 08 Nov 2018 15:14
 Operator : MD/SY
 Sample : J5858-12
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SB-9-S

Quant Time: Nov 08 15:41:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	188057	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	363242	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	319589	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	132726	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	160802	57.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.02%	
35) Dibromofluoromethane	4.89	113	117105	51.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.26%	
50) Toluene-d8	7.57	98	478942	54.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.26%	
62) 4-Bromofluorobenzene	10.31	95	152185	46.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.28%	
Target Compounds						
					Qvalue	
16) Acetone	2.32	43	16337	9.418	ug/l	94
20) Methylene Chloride	2.70	84	9793	3.866	ug/l	99
39) Methylcyclohexane	6.41	83	5008	1.091	ug/l	92
43) Isopropyl Acetate	5.55	43	46775	5.356	ug/l #	95
95) Naphthalene	13.74	128	22166	1.861	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110818\
Data File : VU028026.D
Acq On : 08 Nov 2018 15:14
Operator : MD/SY
Sample : J5858-12
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SB-9-S

Quant Time: Nov 08 15:41:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

