

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU110818\
 Data File : VU028031.D
 Acq On : 08 Nov 2018 17:10
 Operator : MD/SY
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 09 09:27:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	110	0.00
2 T	Dichlorodifluoromethane	0.635	0.495	22.0#	75	0.00
3 P	Chloromethane	0.768	0.658	14.3	87	0.00
4 C	Vinyl Chloride	0.763	0.679	11.0#	91	0.00
5 T	Bromomethane	0.336	0.307	8.6	107	0.00
6 T	Chloroethane	0.439	0.412	6.2	96	0.00
7 T	Trichlorofluoromethane	0.881	0.828	6.0	95	0.00
8 T	Diethyl Ether	0.377	0.370	1.9	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.524	0.525	-0.2	101	0.00
10 T	Methyl Iodide	0.392	0.443	-13.0	108	0.00
11 T	Tert butyl alcohol	0.197	0.187	5.1	99	0.00
12 CM	1,1-Dichloroethene	0.524	0.482	8.0#	96	0.00
13 T	Acrolein	0.078	0.036	53.8#	55	0.00
14 T	Allyl chloride	1.281	1.160	9.4	97	0.00
15 T	Acrylonitrile	0.483	0.482	0.2	100	0.00
16 T	Acetone	0.461	0.454	1.5	107	0.00
17 T	Carbon Disulfide	1.713	1.346	21.4#	81	0.00
18 T	Methyl Acetate	1.269	1.275	-0.5	107	0.00
19 T	Methyl tert-butyl Ether	2.113	2.091	1.0	102	0.00
20 T	Methylene Chloride	0.673	0.643	4.5	99	0.00
21 T	trans-1,2-Dichloroethene	0.605	0.532	12.1	91	0.00
22 T	Diisopropyl ether	2.644	2.611	1.2	101	0.00
23 T	Vinyl Acetate	2.272	2.238	1.5	98	0.00
24 P	1,1-Dichloroethane	1.317	1.256	4.6	99	0.00
25 T	2-Butanone	0.734	0.728	0.8	100	0.00
26 T	2,2-Dichloropropane	0.988	0.977	1.1	106	0.00
27 T	cis-1,2-Dichloroethene	0.678	0.667	1.6	98	0.00
28 T	Bromochloromethane	0.617	0.625	-1.3	110	0.00
29 T	Tetrahydrofuran	0.490	0.461	5.9	97	0.00
30 C	Chloroform	1.208	1.160	4.0#	101	0.00
31 T	Cyclohexane	1.547	1.155	25.3#	94	0.00
32 T	1,1,1-Trichloroethane	0.961	0.927	3.5	98	0.00
33 S	1,2-Dichloroethane-d4	0.743	0.714	3.9	106	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	110	0.00
35 S	Dibromofluoromethane	0.315	0.307	2.5	108	0.00
36 T	1,1-Dichloropropene	0.524	0.482	8.0	97	0.00
37 T	Ethyl Acetate	0.767	0.733	4.4	99	0.00
38 T	Carbon Tetrachloride	0.464	0.435	6.3	96	0.00
39 T	Methylcyclohexane	0.632	0.589	6.8	97	0.00
40 TM	Benzene	1.601	1.526	4.7	97	0.00
41 T	Methacrylonitrile	0.409	0.417	-2.0	99	0.00
42 TM	1,2-Dichloroethane	0.569	0.541	4.9	98	0.00
43 T	Isopropyl Acetate	1.202	1.185	1.4	100	0.00
44 TM	Trichloroethene	0.344	0.331	3.8	99	0.00
45 C	1,2-Dichloropropane	0.462	0.457	1.1#	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU110818\
 Data File : VU028031.D
 Acq On : 08 Nov 2018 17:10
 Operator : MD/SY
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 09 09:27:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.265	0.262	1.1	101	0.00
47 T	Bromodichloromethane	0.511	0.521	-2.0	102	0.00
48 T	Methyl methacrylate	0.589	0.594	-0.8	100	0.00
49 T	1,4-Dioxane	0.010	0.010	0.0	103	0.00
50 S	Toluene-d8	1.218	1.174	3.6	106	0.00
51 T	4-Methyl-2-Pentanone	0.803	0.791	1.5	100	0.00
52 CM	Toluene	0.941	0.901	4.3#	98	0.00
53 T	t-1,3-Dichloropropene	0.572	0.608	-6.3	102	0.00
54 T	cis-1,3-Dichloropropene	0.631	0.652	-3.3	101	0.00
55 T	1,1,2-Trichloroethane	0.383	0.383	0.0	105	0.00
56 T	Ethyl methacrylate	0.675	0.697	-3.3	102	0.00
57 T	1,3-Dichloropropane	0.698	0.703	-0.7	103	0.00
58 T	2-Chloroethyl Vinyl ether	0.362	0.367	-1.4	108	0.00
59 T	2-Hexanone	0.621	0.620	0.2	100	0.00
60 T	Dibromochloromethane	0.362	0.377	-4.1	102	0.00
61 T	1,2-Dibromoethane	0.407	0.404	0.7	99	0.00
62 S	4-Bromofluorobenzene	0.449	0.453	-0.9	111	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	110	0.00
64 T	Tetrachloroethene	0.324	0.288	11.1	97	0.00
65 PM	Chlorobenzene	1.056	1.023	3.1	102	0.00
66 T	1,1,1,2-Tetrachloroethane	0.361	0.358	0.8	102	0.00
67 C	Ethyl Benzene	1.944	1.887	2.9#	101	0.00
68 T	m/p-Xylenes	0.706	0.689	2.4	101	0.00
69 T	o-Xylene	0.712	0.682	4.2	101	0.00
70 T	Styrene	1.138	1.128	0.9	101	0.00
71 P	Bromoform	0.302	0.308	-2.0	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
73 T	Isopropylbenzene	3.873	3.737	3.5	103	0.00
74 T	N-amyl acetate	2.342	2.290	2.2	100	0.00
75 P	1,1,2,2-Tetrachloroethane	1.562	1.501	3.9	103	0.00
76 T	1,2,3-Trichloropropane	1.368	1.508	-10.2	121	0.00
77 T	Bromobenzene	0.888	0.869	2.1	101	0.00
78 T	n-propylbenzene	4.636	4.546	1.9	103	0.00
79 T	2-Chlorotoluene	2.779	2.621	5.7	102	0.00
80 T	1,3,5-Trimethylbenzene	3.209	3.114	3.0	101	0.00
81 T	trans-1,4-Dichloro-2-butene	0.507	0.588	-16.0	104	0.00
82 T	4-Chlorotoluene	3.155	3.034	3.8	103	0.00
83 T	tert-Butylbenzene	3.208	3.072	4.2	103	0.00
84 T	1,2,4-Trimethylbenzene	3.231	3.215	0.5	103	0.00
85 T	sec-Butylbenzene	3.901	3.859	1.1	102	0.00
86 T	p-Isopropyltoluene	3.261	3.231	0.9	103	0.00
87 T	1,3-Dichlorobenzene	1.641	1.578	3.8	102	0.00
88 T	1,4-Dichlorobenzene	1.682	1.590	5.5	102	0.00
89 T	n-Butylbenzene	3.099	3.162	-2.0	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU110818\
Data File : VU028031.D
Acq On : 08 Nov 2018 17:10
Operator : MD/SY
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC050

Quant Time: Nov 09 09:27:08 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.581	0.573	1.4	103	0.00
91 T	1,2-Dichlorobenzene	1.650	1.614	2.2	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.329	0.330	-0.3	102	0.00
93 T	1,2,4-Trichlorobenzene	1.041	1.098	-5.5	102	0.00
94 T	Hexachlorobutadiene	0.524	0.524	0.0	104	0.00
95 T	Naphthalene	3.725	3.986	-7.0	101	0.00
96 T	1,2,3-Trichlorobenzene	1.107	1.117	-0.9	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6