

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110823\
 Data File : VU056124.D
 Acq On : 08 Nov 2023 15:16
 Operator : MD/SY
 Sample : 05297-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B24

Quant Time: Nov 08 20:52:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 08 20:49:31 2023
 Response via : Initial Calibration

Compound			R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards								
1) 1,4-Difluorobenzene			6.245	114	362535	5.000	ug/L	0.00
28) Chlorobenzene-d5			9.415	117	361663	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4			11.807	152	159981	5.000	ug/L	0.00
System Monitoring Compounds								
4) Vinyl Chloride-d3			1.595	65	112805	5.013	ug/L	0.00
Spiked Amount 5.000			Range 40 - 130		Recovery	= 100.200%		
7) Chloroethane-d5			1.907	69	89129	5.837	ug/L	0.00
Spiked Amount 5.000			Range 65 - 130		Recovery	= 116.800%		
11) 1,1-Dichloroethene-d2			2.563	65	50442	5.088	ug/L	0.00
Spiked Amount 5.000			Range 60 - 125		Recovery	= 101.800%		
20) 2-Butanone-d5			4.618	46	232503	62.981	ug/L	0.00
Spiked Amount 50.000			Range 40 - 130		Recovery	= 125.960%		
24) Chloroform-d			5.055	84	227241	5.305	ug/L	0.00
Spiked Amount 5.000			Range 70 - 125		Recovery	= 106.200%		
26) 1,2-Dichloroethane-d4			5.698	65	106464	5.645	ug/L	0.00
Spiked Amount 5.000			Range 70 - 130		Recovery	= 112.800%		
32) Benzene-d6			5.721	84	451675	5.200	ug/L	0.00
Spiked Amount 5.000			Range 70 - 125		Recovery	= 104.000%		
36) 1,2-Dichloropropane-d6			6.685	67	141028	5.516	ug/L	0.00
Spiked Amount 5.000			Range 60 - 140		Recovery	= 110.400%		
41) Toluene-d8			7.894	98	386436	4.855	ug/L	0.00
Spiked Amount 5.000			Range 70 - 130		Recovery	= 97.000%		
43) trans-1,3-Dichloroprop...			8.177	79	51168	5.235	ug/L	0.00
Spiked Amount 5.000			Range 55 - 130		Recovery	= 104.800%		
46) 2-Hexanone-d5			8.631	63	178900	55.846	ug/L	0.00
Spiked Amount 50.000			Range 45 - 130		Recovery	= 111.700%		
56) 1,1,2,2-Tetrachloroeth...			10.753	84	103901	5.693	ug/L	0.00
Spiked Amount 5.000			Range 65 - 120		Recovery	= 113.800%		
66) 1,2-Dichlorobenzene-d4			12.190	152	138656	5.620	ug/L	0.00
Spiked Amount 5.000			Range 80 - 120		Recovery	= 112.400%		
Target Compounds								Qvalue
13) Acetone			2.628	43	38433	14.116	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110823\
Data File : VU056124.D
Acq On : 08 Nov 2023 15:16
Operator : MD/SY
Sample : 05297-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0B24

Quant Time: Nov 08 20:52:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 08 20:49:31 2023
Response via : Initial Calibration

