

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110823\
 Data File : VU056113.D
 Acq On : 08 Nov 2023 10:35
 Operator : MD/SY
 Sample : 05281-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AD2

Quant Time: Nov 08 20:50:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 08 20:49:31 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	448957	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	439664	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.810	152	190577	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	102698	3.685	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.800%
7) Chloroethane-d5	1.907	69	84960	4.493	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.800%
11) 1,1-Dichloroethene-d2	2.563	65	45000	3.665	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.400%
20) 2-Butanone-d5	4.621	46	186414	40.776	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	81.560%
24) Chloroform-d	5.058	84	205410	3.872	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.400%
26) 1,2-Dichloroethane-d4	5.695	65	94754	4.057	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.200%
32) Benzene-d6	5.724	84	413845	3.919	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.400%
36) 1,2-Dichloropropane-d6	6.685	67	124029	3.990	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	79.800%
41) Toluene-d8	7.894	98	351727	3.635	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.600%
43) trans-1,3-Dichloroprop...	8.177	79	44963	3.784	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.600%
46) 2-Hexanone-d5	8.630	63	142641	36.628	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	73.260%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	87648	3.951	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.000%
66) 1,2-Dichlorobenzene-d4	12.190	152	124287	4.229	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	84.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110823\
Data File : VU056113.D
Acq On : 08 Nov 2023 10:35
Operator : MD/SY
Sample : 05281-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AD2

Quant Time: Nov 08 20:50:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 08 20:49:31 2023
Response via : Initial Calibration

