

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110823\
 Data File : VU056115.D
 Acq On : 08 Nov 2023 11:37
 Operator : MD/SY
 Sample : 05281-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AC5

Quant Time: Nov 08 20:50:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 08 20:49:31 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	362500	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	355635	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	155562	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	118851	5.282	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 105.600%		
7) Chloroethane-d5	1.907	69	98603	6.458	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 129.200%		
11) 1,1-Dichloroethene-d2	2.563	65	53294	5.376	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 107.600%		
20) 2-Butanone-d5	4.618	46	241179	65.337	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 130.680%#		
24) Chloroform-d	5.058	84	238337	5.565	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 111.200%		
26) 1,2-Dichloroethane-d4	5.695	65	113861	6.037	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 120.800%		
32) Benzene-d6	5.721	84	484707	5.675	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 113.400%		
36) 1,2-Dichloropropane-d6	6.685	67	145228	5.776	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 115.600%		
41) Toluene-d8	7.894	98	411839	5.262	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 105.200%		
43) trans-1,3-Dichloroprop...	8.177	79	53988	5.617	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 112.400%		
46) 2-Hexanone-d5	8.627	63	184310	58.510	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 117.020%		
56) 1,1,2,2-Tetrachloroeth...	10.753	84	104285	5.811	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 116.200%		
66) 1,2-Dichlorobenzene-d4	12.190	152	145365	6.060	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 121.200%#		
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.576	96	498604	21.401	ug/L #	83
19) 1,1-Dichloroethane	3.862	63	127811	2.838	ug/L	96
22) cis-1,2-Dichloroethene	4.660	96	8505	0.319	ug/L	93
29) 1,1,1-Trichloroethane	5.309	97	189511	4.651	ug/L	99
34) Trichloroethene	6.541	95	18748	0.674	ug/L	97
47) Tetrachloroethene	8.550	164	15338	0.735	ug/L	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110823\
Data File : VU056115.D
Acq On : 08 Nov 2023 11:37
Operator : MD/SY
Sample : 05281-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AC5

Quant Time: Nov 08 20:50:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 08 20:49:31 2023
Response via : Initial Calibration





#12

1,1-Dichloroethene

Concen: 21.401 ug/L

RT: 2.576 min Scan# 400

Delta R.T. -0.000 min

Lab File: VU056115.D

Acq: 08 Nov 2023 11:37

Instrument :

MSVOA_U

ClientSampleId :

C0AC5





Tgt Ion: 96 Resp: 498604

Ion Ratio Lower Upper

96 100

61 157.9 111.7 207.5

63 65.1 74.5 138.3#



#19

1,1-Dichloroethane

Concen: 2.838 ug/L

RT: 3.862 min Scan# 800

Delta R.T. -0.000 min

Lab File: VU056115.D

Acq: 08 Nov 2023 11:37




Tgt Ion: 63 Resp: 127811


Ion Ratio Lower Upper

63 100

65 31.8 20.6 38.2

83 14.9 9.6 17.8





#22

cis-1,2-Dichloroethene

Concen: 0.319 ug/L

RT: 4.660 min Scan# 1048

Delta R.T. -0.000 min

Lab File: VU056115.D

Acq: 08 Nov 2023 11:37

Instrument :

MSVOA_U

ClientSampleId :

C0AC5





Tgt Ion: 96 Resp: 8505

Ion Ratio Lower Upper

96 100

61 123.5 87.8 163.2

98 48.3 43.5 80.7



#29

1,1,1-Trichloroethane

Concen: 4.651 ug/L

RT: 5.309 min Scan# 1250

Delta R.T. -0.000 min

Lab File: VU056115.D

Acq: 08 Nov 2023 11:37




Tgt Ion: 97 Resp: 189511


Ion Ratio Lower Upper

97 100

99 65.2 52.0 78.0

61 41.1 33.4 50.0









#34
Trichloroethene
Concen: 0.674 ug/L
RT: 6.541 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU056115.D
Acq: 08 Nov 2023 11:37

Instrument :
MSVOA_U
ClientSampleId :
C0AC5

Tgt Ion: 95 Resp: 18748
Ion Ratio Lower Upper
95 100
97 67.0 46.8 87.0
132 88.3 66.5 123.5
130 101.3 70.6 131.0



#47
Tetrachloroethene
Concen: 0.735 ug/L
RT: 8.550 min Scan# 2258
Delta R.T. -0.000 min
Lab File: VU056115.D
Acq: 08 Nov 2023 11:37

Tgt Ion: 164 Resp: 15338
Ion Ratio Lower Upper
164 100
129 103.5 66.6 123.8
131 98.0 67.8 126.0
166 128.1 85.9 159.5

