

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110823\
 Data File : VU056123.D
 Acq On : 08 Nov 2023 14:52
 Operator : MD/SY
 Sample : 05281-09
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AD0

Quant Time: Nov 08 20:52:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 08 20:49:31 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	367757	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	371427	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	159340	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	121337	5.316	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.400%
7) Chloroethane-d5	1.907	69	92571	5.976	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	119.600%
11) 1,1-Dichloroethene-d2	2.560	65	52623	5.233	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	104.600%
20) 2-Butanone-d5	4.621	46	242440	64.740	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	129.480%
24) Chloroform-d	5.055	84	237655	5.469	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.400%
26) 1,2-Dichloroethane-d4	5.695	65	114266	5.972	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	119.400%
32) Benzene-d6	5.721	84	482389	5.408	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.200%
36) 1,2-Dichloropropane-d6	6.685	67	148680	5.662	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	113.200%
41) Toluene-d8	7.894	98	416066	5.090	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.800%
43) trans-1,3-Dichloroprop...	8.177	79	52845	5.265	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	105.200%
46) 2-Hexanone-d5	8.627	63	189744	57.674	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	115.340%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	108280	5.777	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	115.600%
66) 1,2-Dichlorobenzene-d4	12.190	152	141894	5.775	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110823\
Data File : VU056123.D
Acq On : 08 Nov 2023 14:52
Operator : MD/SY
Sample : 05281-09
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AD0

Quant Time: Nov 08 20:52:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 08 20:49:31 2023
Response via : Initial Calibration

