

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110918\
 Data File : VU028039.D
 Acq On : 09 Nov 2018 13:04
 Operator : MD/SY
 Sample : J5885-11
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TRIP-BLANK

Quant Time: Nov 12 11:36:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

| Internal Standards | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) Pentafluorobenzene | 4.99 | 168 | 195949 | 50.00 | ug/l | 0.00 |
| 34) 1,4-Difluorobenzene | 5.89 | 114 | 377764 | 50.00 | ug/l | 0.00 |
| 63) Chlorobenzene-d5 | 9.09 | 117 | 338167 | 50.00 | ug/l | 0.00 |
| 72) 1,4-Dichlorobenzene-d4 | 11.48 | 152 | 142309 | 50.00 | ug/l | 0.00 |

System Monitoring Compounds

| | | | | | | |
|---------------------------|-------|-----|--------|----------|------|---------|
| 33) 1,2-Dichloroethane-d4 | 5.31 | 65 | 169689 | 58.25 | ug/l | 0.00 |
| Spiked Amount | | | | | | |
| | | | | Recovery | = | 116.50% |
| 35) Dibromofluoromethane | 4.88 | 113 | 122976 | 51.63 | ug/l | 0.00 |
| Spiked Amount | | | | | | |
| | | | | Recovery | = | 103.26% |
| 50) Toluene-d8 | 7.57 | 98 | 490428 | 53.30 | ug/l | 0.00 |
| Spiked Amount | | | | | | |
| | | | | Recovery | = | 106.60% |
| 62) 4-Bromofluorobenzene | 10.31 | 95 | 162329 | 47.83 | ug/l | 0.00 |
| Spiked Amount | | | | | | |
| | | | | Recovery | = | 95.66% |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110918\
Data File : VU028039.D
Acq On : 09 Nov 2018 13:04
Operator : MD/SY
Sample : J5885-11
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TRIP-BLANK

Quant Time: Nov 12 11:36:40 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

