

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110918\
 Data File : VU028040.D
 Acq On : 09 Nov 2018 13:27
 Operator : MD/SY
 Sample : J5885-10
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FB11818

Quant Time: Nov 12 11:37:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	196100	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	375347	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	334259	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	131430	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	167881	57.58	ug/l	0.00
Spiked Amount						
			Recovery	=	115.16%	
35) Dibromofluoromethane	4.89	113	120650	50.98	ug/l	0.00
Spiked Amount						
			Recovery	=	101.96%	
50) Toluene-d8	7.57	98	487475	53.32	ug/l	0.00
Spiked Amount						
			Recovery	=	106.64%	
62) 4-Bromofluorobenzene	10.31	95	157296	46.65	ug/l	0.00
Spiked Amount						
			Recovery	=	93.30%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110918\
Data File : VU028040.D
Acq On : 09 Nov 2018 13:27
Operator : MD/SY
Sample : J5885-10
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FB11818

Quant Time: Nov 12 11:37:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

