

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110918\
 Data File : VU028046.D
 Acq On : 09 Nov 2018 15:46
 Operator : MD/SY
 Sample : J5885-05
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-3

Quant Time: Nov 12 07:10:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	194427	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	380818	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	334841	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	143185	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	167757	58.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.06%	
35) Dibromofluoromethane	4.89	113	124325	51.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.56%	
50) Toluene-d8	7.57	98	494938	53.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.72%	
62) 4-Bromofluorobenzene	10.31	95	164068	47.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.92%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.40	62	57895	19.523	ug/l	99
16) Acetone	2.32	43	3654	2.037	ug/l #	80
21) trans-1,2-Dichloroethene	2.99	96	9840	4.184	ug/l	92
27) cis-1,2-Dichloroethene	4.23	96	879851	333.650	ug/l	90
44) Trichloroethene	6.19	130	341901	130.438	ug/l	100
64) Tetrachloroethene	8.23	164	5444762	2507.721	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110918\
Data File : VU028046.D
Acq On : 09 Nov 2018 15:46
Operator : MD/SY
Sample : J5885-05
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-3

Quant Time: Nov 12 07:10:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

