

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110918\
 Data File : VU028049.D
 Acq On : 09 Nov 2018 16:57
 Operator : MD/SY
 Sample : J5885-08
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-9

Quant Time: Nov 12 07:22:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	195912	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	380848	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	342586	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	137940	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	167719	57.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.16%	
35) Dibromofluoromethane	4.88	113	127382	53.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.10%	
50) Toluene-d8	7.57	98	501630	54.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.16%	
62) 4-Bromofluorobenzene	10.31	95	162643	47.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.08%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.41	62	5395	1.805	ug/l	95
16) Acetone	2.32	43	5945	3.290	ug/l	98
27) cis-1,2-Dichloroethene	4.23	96	20291	7.636	ug/l	93
44) Trichloroethene	6.19	130	4803	1.832	ug/l	93
64) Tetrachloroethene	8.23	164	346233	155.861	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110918\
Data File : VU028049.D
Acq On : 09 Nov 2018 16:57
Operator : MD/SY
Sample : J5885-08
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-9

Quant Time: Nov 12 07:22:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

