

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110918\
 Data File : VU028052.D
 Acq On : 09 Nov 2018 18:09
 Operator : MD/SY
 Sample : I.BLK
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 I.BLK

Quant Time: Nov 12 07:27:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 02 02:28:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	195460	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	376827	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	345471	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	140995	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	168312	57.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.84%	
35) Dibromofluoromethane	4.88	113	126130	53.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.18%	
50) Toluene-d8	7.57	98	499097	54.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.76%	
62) 4-Bromofluorobenzene	10.31	95	164939	48.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.44%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.23	96	4996	1.885	ug/l	79
44) Trichloroethene	6.19	130	1139	0.439	ug/l	55
64) Tetrachloroethene	8.23	164	57002	25.446	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110918\
Data File : VU028052.D
Acq On : 09 Nov 2018 18:09
Operator : MD/SY
Sample : I.BLK
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
I.BLK

Quant Time: Nov 12 07:27:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U110218W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 02 02:28:20 2018
Response via : Initial Calibration

