

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U110920\
 Data File : VU041134.D
 Acq On : 09 Nov 2020 12:37
 Operator : SY/MD
 Sample : VU1109WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK18

Quant Time: Nov 10 01:19:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Nov 10 01:04:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	55950	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	58768	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	25508	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	24919	5.30	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.00%
7) Chloroethane-d5	1.92	69	19295	5.34	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.80%
11) 1,1-Dichloroethene-d2	2.58	63	35686	3.87	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.40%
20) 2-Butanone-d5	4.65	46	79564	51.40	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.80%
24) Chloroform-d	5.08	84	44946	5.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.20%
26) 1,2-Dichloroethane-d4	5.72	65	28286	5.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.80%
32) Benzene-d6	5.74	84	83005	5.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.60%
36) 1,2-Dichloropropane-d6	6.70	67	27824	5.44	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.80%
41) Toluene-d8	7.90	98	68891	4.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.80%
43) trans-1,3-Dichloropropene-	8.19	79	11283	5.09	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	101.80%
46) 2-Hexanone-d5	8.64	63	56157	47.02	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.04%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	23214	4.88	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	26669	5.96	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	119.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU110920\
Data File : VU041134.D
Acq On : 09 Nov 2020 12:37
Operator : SY/MD
Sample : VU1109WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK18

Quant Time: Nov 10 01:19:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Nov 10 01:04:12 2020
Response via : Initial Calibration

