

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110920\
 Data File : VU041140.D
 Acq On : 09 Nov 2020 15:03
 Operator : SY/MD
 Sample : L4646-24
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GB0W8

Quant Time: Nov 10 01:36:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Nov 10 01:04:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	54472	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	57031	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	23874	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	16217	3.54	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	70.80%
7) Chloroethane-d5	1.92	69	14706	4.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.60%
11) 1,1-Dichloroethene-d2	2.58	63	25896	2.88	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	57.60%#
20) 2-Butanone-d5	4.65	46	86498	57.40	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.80%
24) Chloroform-d	5.08	84	42540	5.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.20%
26) 1,2-Dichloroethane-d4	5.72	65	27728	5.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.60%
32) Benzene-d6	5.74	84	72835	4.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.60%
36) 1,2-Dichloropropane-d6	6.70	67	26669	5.37	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.40%
41) Toluene-d8	7.91	98	58281	4.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.40%
43) trans-1,3-Dichloropropene-	8.19	79	9864	4.58	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.60%
46) 2-Hexanone-d5	8.64	63	60892	52.54	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.08%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	24442	5.30	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.00%
64) 1,2-Dichlorobenzene-d4	12.19	152	25818	6.16	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	123.20%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU110920\
Data File : VU041140.D
Acq On : 09 Nov 2020 15:03
Operator : SY/MD
Sample : L4646-24
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GB0W8

Quant Time: Nov 10 01:36:55 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Nov 10 01:04:12 2020
Response via : Initial Calibration

