

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110921\
 Data File : VU045650.D
 Acq On : 09 Nov 2021 11:55
 Operator : SY/MD
 Sample : VU1109WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1109WBL01

Quant Time: Nov 10 02:29:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	150171	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	277374	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	270596	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	125925	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	119305	53.048	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.100%
35) Dibromofluoromethane	5.292	113	94391	50.970	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.940%
50) Toluene-d8	7.902	98	362081	52.581	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.160%
62) 4-Bromofluorobenzene	10.636	95	130222	47.159	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110921\
Data File : VU045650.D
Acq On : 09 Nov 2021 11:55
Operator : SY/MD
Sample : VU1109WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1109WBL01

Quant Time: Nov 10 02:29:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

