

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110921\
 Data File : VU045653.D
 Acq On : 09 Nov 2021 13:14
 Operator : SY/MD
 Sample : M4591-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TW-F

Quant Time: Nov 10 01:58:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	144525	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	264630	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	262250	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	128744	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	113416	52.400	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.800%	
35) Dibromofluoromethane	5.292	113	89716	50.779	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.560%	
50) Toluene-d8	7.899	98	343454	52.278	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.560%	
62) 4-Bromofluorobenzene	10.636	95	129635	49.207	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.420%	
Target Compounds						Qvalue
89) n-Butylbenzene	12.211	91	2403	0.321	ug/l	98
93) 1,2,4-Trichlorobenzene	13.844	180	1445	0.521	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110921\
Data File : VU045653.D
Acq On : 09 Nov 2021 13:14
Operator : SY/MD
Sample : M4591-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TW-F

Quant Time: Nov 10 01:58:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

