

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110921\
 Data File : VU045654.D
 Acq On : 09 Nov 2021 13:37
 Operator : SY/MD
 Sample : M4591-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TW-H

Quant Time: Nov 10 01:58:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	146959	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	275061	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	277997	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	138358	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	119578	54.332	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.660%	
35) Dibromofluoromethane	5.292	113	92924	50.600	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.200%	
50) Toluene-d8	7.899	98	358446	52.491	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.980%	
62) 4-Bromofluorobenzene	10.636	95	137240	50.119	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.240%	
Target Compounds						
					Qvalue	
16) Acetone	2.626	43	7886	7.039	ug/l	99
83) tert-Butylbenzene	11.423	119	2200	0.292	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110921\
Data File : VU045654.D
Acq On : 09 Nov 2021 13:37
Operator : SY/MD
Sample : M4591-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TW-H

Quant Time: Nov 10 01:58:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

