

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110921\
 Data File : VU045662.D
 Acq On : 09 Nov 2021 16:42
 Operator : SY/MD
 Sample : M4594-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 08BR-20211108

Quant Time: Nov 10 02:00:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	137164	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	262409	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	264680	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	129892	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	113825	55.411	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.820%
35) Dibromofluoromethane	5.292	113	88679	50.617	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.240%
50) Toluene-d8	7.899	98	343796	52.773	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.540%
62) 4-Bromofluorobenzene	10.636	95	130616	49.999	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.000%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.585	101	10875	7.105	ug/l	99
16) Acetone	2.626	43	2435	2.329	ug/l #	86
27) cis-1,2-Dichloroethene	4.671	96	21378	11.594	ug/l	90
44) Trichloroethene	6.546	130	438664	241.915	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110921\
Data File : VU045662.D
Acq On : 09 Nov 2021 16:42
Operator : SY/MD
Sample : M4594-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
08BR-20211108

Quant Time: Nov 10 02:00:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

