

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110921\
 Data File : VU045668.D
 Acq On : 09 Nov 2021 19:00
 Operator : SY/MD
 Sample : M4594-07
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 48BR-20211108

Quant Time: Nov 10 02:01:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	137793	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	259604	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	264444	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	133837	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	112033	54.290	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.580%	
35) Dibromofluoromethane	5.292	113	87143	50.278	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.560%	
50) Toluene-d8	7.899	98	345556	53.616	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.240%	
62) 4-Bromofluorobenzene	10.636	95	132203	51.154	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.300%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.584	96	6892	4.546	ug/l	96
16) Acetone	2.629	43	2628	2.502	ug/l	93
27) cis-1,2-Dichloroethene	4.671	96	40545	21.888	ug/l	91
44) Trichloroethene	6.546	130	1106250	616.668	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110921\
Data File : VU045668.D
Acq On : 09 Nov 2021 19:00
Operator : SY/MD
Sample : M4594-07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
48BR-20211108

Quant Time: Nov 10 02:01:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

