

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110921\
 Data File : VU045671.D
 Acq On : 09 Nov 2021 20:10
 Operator : SY/MD
 Sample : M4594-12
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BRP2-20211108

Quant Time: Nov 10 02:02:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	131525	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	245048	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	247074	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	121424	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	107539	54.596	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.200%
35) Dibromofluoromethane	5.292	113	83086	50.784	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.560%
50) Toluene-d8	7.899	98	324474	53.335	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.680%
62) 4-Bromofluorobenzene	10.635	95	121408	49.767	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.540%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.604	62	5137	2.703	ug/l	95
16) Acetone	2.629	43	2469	2.462	ug/l	97
27) cis-1,2-Dichloroethene	4.671	96	50032	28.297	ug/l	92
44) Trichloroethene	6.546	130	33032	19.507	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110921\
Data File : VU045671.D
Acq On : 09 Nov 2021 20:10
Operator : SY/MD
Sample : M4594-12
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BRP2-20211108

Quant Time: Nov 10 02:02:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

