

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110921\
 Data File : VU045672.D
 Acq On : 09 Nov 2021 20:33
 Operator : SY/MD
 Sample : M4594-14
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TB01-20211108

Quant Time: Nov 10 02:02:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	130585	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	248484	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	256555	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	127388	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	109682	56.084	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.160%
35) Dibromofluoromethane	5.292	113	86253	51.991	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.980%
50) Toluene-d8	7.902	98	328066	53.180	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.360%
62) 4-Bromofluorobenzene	10.636	95	126210	51.020	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.040%
Target Compounds						
						Qvalue
16) Acetone	2.629	43	2225	2.235	ug/l	# 85
27) cis-1,2-Dichloroethene	4.671	96	4874m	2.776	ug/l	
44) Trichloroethene	6.546	130	10576	6.159	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110921\
Data File : VU045672.D
Acq On : 09 Nov 2021 20:33
Operator : SY/MD
Sample : M4594-14
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TB01-20211108

Quant Time: Nov 10 02:02:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

