

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110921\
 Data File : VU045673.D
 Acq On : 09 Nov 2021 20:56
 Operator : SY/MD
 Sample : M4594-15
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FB01-20211108

Quant Time: Nov 10 02:03:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	132624	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	247067	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	253711	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	128806	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	109590	55.176	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 110.360%	
35) Dibromofluoromethane	5.292	113	85655	51.927	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.860%	
50) Toluene-d8	7.899	98	326326	53.201	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.400%	
62) 4-Bromofluorobenzene	10.636	95	125306	50.945	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.900%	
Target Compounds						
					Qvalue	
16) Acetone	2.626	43	4473	4.424	ug/l	98
27) cis-1,2-Dichloroethene	4.671	96	2949	1.654	ug/l	93
44) Trichloroethene	6.546	130	8009	4.691	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110921\
Data File : VU045673.D
Acq On : 09 Nov 2021 20:56
Operator : SY/MD
Sample : M4594-15
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FB01-20211108

Quant Time: Nov 10 02:03:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

