

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110921\
 Data File : VU045677.D
 Acq On : 09 Nov 2021 23:14
 Operator : SY/MD
 Sample : VU1109WBL02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1109WBL02

Quant Time: Nov 10 02:04:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.376	168	129858	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	243251	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	247075	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	125137	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	107545	55.300	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.600%
35) Dibromofluoromethane	5.292	113	83080	51.156	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.320%
50) Toluene-d8	7.900	98	319852	52.964	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.920%
62) 4-Bromofluorobenzene	10.636	95	121452	50.153	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110921\
Data File : VU045677.D
Acq On : 09 Nov 2021 23:14
Operator : SY/MD
Sample : VU1109WBL02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1109WBL02

Quant Time: Nov 10 02:04:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

