

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110921\
 Data File : VU045648.D
 Acq On : 09 Nov 2021 10:19
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 10 01:57:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	107	0.00
2 T	Dichlorodifluoromethane	0.648	0.624	3.7	100	0.00
3 P	Chloromethane	0.742	0.658	11.3	97	0.00
4 C	Vinyl Chloride	0.723	0.707	2.2#	102	0.00
5 T	Bromomethane	0.403	0.429	-6.5	104	0.00
6 T	Chloroethane	0.472	0.425	10.0	100	0.00
7 T	Trichlorofluoromethane	0.955	0.923	3.4	103	0.00
8 T	Diethyl Ether	0.378	0.375	0.8	104	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.558	0.547	2.0	104	0.00
10 T	Methyl Iodide	0.730	0.603	17.4	85	0.00
11 T	Tert butyl alcohol	0.148	0.140	5.4	96	0.00
12 CM	1,1-Dichloroethene	0.550	0.521	5.3#	101	0.00
13 T	Acrolein	0.021	0.019	9.5	100	0.00
14 T	Allyl chloride	0.982	0.941	4.2	100	0.00
15 T	Acrylonitrile	0.396	0.381	3.8	99	0.00
16 T	Acetone	0.381	0.405	-6.3	107	0.00
17 T	Carbon Disulfide	1.541	1.366	11.4	96	0.00
18 T	Methyl Acetate	1.298	1.219	6.1	98	0.00
19 T	Methyl tert-butyl Ether	1.980	1.959	1.1	103	0.00
20 T	Methylene Chloride	0.662	0.626	5.4	104	0.00
21 T	trans-1,2-Dichloroethene	0.572	0.572	0.0	104	0.00
22 T	Diisopropyl ether	1.929	1.922	0.4	102	0.00
23 T	Vinyl Acetate	1.483	1.499	-1.1	102	0.00
24 P	1,1-Dichloroethane	1.149	1.124	2.2	102	0.00
25 T	2-Butanone	0.529	0.538	-1.7	102	0.00
26 T	2,2-Dichloropropane	0.952	0.958	-0.6	102	0.00
27 T	cis-1,2-Dichloroethene	0.672	0.666	0.9	104	0.00
28 T	Bromochloromethane	0.548	0.470	14.2	91	0.00
29 T	Tetrahydrofuran	0.323	0.309	4.3	97	0.00
30 C	Chloroform	1.137	1.127	0.9#	105	0.00
31 T	Cyclohexane	1.093	1.006	8.0	100	0.00
32 T	1,1,1-Trichloroethane	0.969	0.965	0.4	102	0.00
33 S	1,2-Dichloroethane-d4	0.749	0.646	13.8	97	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	106	0.00
35 S	Dibromofluoromethane	0.334	0.289	13.5	97	0.00
36 T	1,1-Dichloropropene	0.488	0.479	1.8	102	0.00
37 T	Ethyl Acetate	0.560	0.540	3.6	99	0.00
38 T	Carbon Tetrachloride	0.492	0.474	3.7	102	0.00
39 T	Methylcyclohexane	0.598	0.592	1.0	101	0.00
40 TM	Benzene	1.443	1.395	3.3	102	0.00
41 T	Methacrylonitrile	0.296	0.301	-1.7	101	0.00
42 TM	1,2-Dichloroethane	0.521	0.519	0.4	103	0.00
43 T	Isopropyl Acetate	0.857	0.841	1.9	98	0.00
44 TM	Trichloroethene	0.346	0.348	-0.6	107	0.00
45 C	1,2-Dichloropropane	0.382	0.382	0.0	103	0.00
46 T	Dibromomethane	0.265	0.262	1.1	104	0.00
47 T	Bromodichloromethane	0.502	0.510	-1.6	103	0.00
48 T	Methyl methacrylate	0.394	0.397	-0.8	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110921\
 Data File : VU045648.D
 Acq On : 09 Nov 2021 10:19
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 10 01:57:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.010	0.0	99	0.00
50 S	Toluene-d8	1.241	1.092	12.0	94	0.00
51 T	4-Methyl-2-Pentanone	0.565	0.559	1.1	99	0.00
52 CM	Toluene	0.884	0.891	-0.8#	102	0.00
53 T	t-1,3-Dichloropropene	0.577	0.584	-1.2	101	0.00
54 T	cis-1,3-Dichloropropene	0.613	0.636	-3.8	104	0.00
55 T	1,1,2-Trichloroethane	0.366	0.370	-1.1	105	0.00
56 T	Ethyl methacrylate	0.583	0.600	-2.9	102	0.00
57 T	1,3-Dichloropropane	0.660	0.651	1.4	103	0.00
58 T	2-Chloroethyl Vinyl ether	0.337	0.277	17.8	88	0.00
59 T	2-Hexanone	0.445	0.444	0.2	98	0.00
60 T	Dibromochloromethane	0.381	0.390	-2.4	105	0.00
61 T	1,2-Dibromoethane	0.393	0.393	0.0	103	0.00
62 S	4-Bromofluorobenzene	0.498	0.430	13.7	97	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	103	0.00
64 T	Tetrachloroethene	0.287	0.298	-3.8	105	0.00
65 PM	Chlorobenzene	0.982	0.981	0.1	103	0.00
66 T	1,1,1,2-Tetrachloroethane	0.350	0.350	0.0	103	0.00
67 C	Ethyl Benzene	1.759	1.816	-3.2#	104	0.00
68 T	m/p-Xylenes	0.663	0.701	-5.7	104	0.00
69 T	o-Xylene	0.645	0.676	-4.8	104	0.00
70 T	Styrene	1.091	1.153	-5.7	103	0.00
71 P	Bromoform	0.292	0.299	-2.4	102	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
73 T	Isopropylbenzene	3.267	3.461	-5.9	105	0.00
74 T	N-amyl acetate	1.474	1.501	-1.8	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.294	1.256	2.9	100	0.00
76 T	1,2,3-Trichloropropane	1.516	1.016	33.0#	68	0.00
77 T	Bromobenzene	0.793	0.806	-1.6	104	0.00
78 T	n-propylbenzene	4.066	4.221	-3.8	104	0.00
79 T	2-Chlorotoluene	2.402	2.459	-2.4	103	0.00
80 T	1,3,5-Trimethylbenzene	2.854	2.981	-4.4	103	0.00
81 T	trans-1,4-Dichloro-2-butene	0.425	0.407	4.2	94	0.00
82 T	4-Chlorotoluene	2.817	2.889	-2.6	102	0.00
83 T	tert-Butylbenzene	2.721	2.845	-4.6	103	0.00
84 T	1,2,4-Trimethylbenzene	2.837	2.999	-5.7	102	0.00
85 T	sec-Butylbenzene	3.609	3.739	-3.6	103	0.00
86 T	p-Isopropyltoluene	3.016	3.144	-4.2	102	0.00
87 T	1,3-Dichlorobenzene	1.600	1.574	1.6	103	0.00
88 T	1,4-Dichlorobenzene	1.619	1.597	1.4	103	0.00
89 T	n-Butylbenzene	2.905	3.005	-3.4	103	0.00
90 T	Hexachloroethane	0.544	0.534	1.8	102	0.00
91 T	1,2-Dichlorobenzene	1.557	1.546	0.7	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.305	0.298	2.3	96	0.00
93 T	1,2,4-Trichlorobenzene	1.077	1.075	0.2	102	0.00
94 T	Hexachlorobutadiene	0.465	0.406	12.7	93	0.00
95 T	Naphthalene	3.671	3.843	-4.7	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110921\
Data File : VU045648.D
Acq On : 09 Nov 2021 10:19
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC050

Quant Time: Nov 10 01:57:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.069	1.086	-1.6	102	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 5