

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110922\
 Data File : VU051791.D
 Acq On : 09 Nov 2022 11:37
 Operator : JC/MD
 Sample : VU1109WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK133

Quant Time: Nov 10 01:04:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 10 01:04:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	224945	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	239135	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	100699	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	74399	4.391	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.800%
7) Chloroethane-d5	1.912	69	70626	3.997	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.000%
11) 1,1-Dichloroethene-d2	2.568	65	31843	5.179	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	103.600%
20) 2-Butanone-d5	4.629	46	184397	30.054	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	60.100%
24) Chloroform-d	5.063	84	149571	3.773	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	75.400%
26) 1,2-Dichloroethane-d4	5.700	65	76740	3.583	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.600%
32) Benzene-d6	5.726	84	304302	4.032	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.600%
36) 1,2-Dichloropropane-d6	6.690	67	98656	3.739	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	74.800%
41) Toluene-d8	7.896	98	240295	4.049	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.000%
43) trans-1,3-Dichloroprop...	8.179	79	30337	3.760	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.200%
46) 2-Hexanone-d5	8.632	63	136024	28.005	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	56.020%
56) 1,1,2,2-Tetrachloroeth...	10.754	84	85766	3.547	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	71.000%
66) 1,2-Dichlorobenzene-d4	12.192	152	91501	4.951	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110922\
Data File : VU051791.D
Acq On : 09 Nov 2022 11:37
Operator : JC/MD
Sample : VU1109WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK133

Quant Time: Nov 10 01:04:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 10 01:04:23 2022
Response via : Initial Calibration

