

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110922\
 Data File : VU051792.D
 Acq On : 09 Nov 2022 12:05
 Operator : JC/MD
 Sample : N5493-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H0AC8

Quant Time: Nov 10 01:05:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 10 01:04:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	206789	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	218259	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	89874	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	71224	4.573	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.400%
7) Chloroethane-d5	1.912	69	64878	3.994	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.800%
11) 1,1-Dichloroethene-d2	2.565	65	29669	5.249	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	105.000%
20) 2-Butanone-d5	4.620	46	174519	30.941	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	61.880%
24) Chloroform-d	5.063	84	139350	3.824	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.400%
26) 1,2-Dichloroethane-d4	5.700	65	73064	3.711	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.200%
32) Benzene-d6	5.726	84	281907	4.093	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.800%
36) 1,2-Dichloropropane-d6	6.690	67	94188	3.911	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	78.200%
41) Toluene-d8	7.896	98	226562	4.183	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.600%
43) trans-1,3-Dichloroprop...	8.179	79	28939	3.930	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.600%
46) 2-Hexanone-d5	8.632	63	132397	29.866	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	59.740%
56) 1,1,2,2-Tetrachloroeth...	10.754	84	83923	3.802	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	76.000%
66) 1,2-Dichlorobenzene-d4	12.192	152	87007	5.275	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110922\
Data File : VU051792.D
Acq On : 09 Nov 2022 12:05
Operator : JC/MD
Sample : N5493-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H0AC8

Quant Time: Nov 10 01:05:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 10 01:04:23 2022
Response via : Initial Calibration

