

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110923\
 Data File : VU056147.D
 Acq On : 09 Nov 2023 18:03
 Operator : MD/SY
 Sample : 05328-07
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 A49W7

Quant Time: Nov 09 21:37:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 09 21:34:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	362829	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	360217	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	157746	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	106038	4.708	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.200%
7) Chloroethane-d5	1.907	69	91574	5.992	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	119.800%
11) 1,1-Dichloroethene-d2	2.563	65	46828	4.720	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	94.400%
20) 2-Butanone-d5	4.618	46	229888	62.222	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	124.440%
24) Chloroform-d	5.058	84	230750	5.383	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.600%
26) 1,2-Dichloroethane-d4	5.698	65	107077	5.673	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.400%
32) Benzene-d6	5.721	84	458416	5.299	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.000%
36) 1,2-Dichloropropane-d6	6.685	67	140788	5.528	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.600%
41) Toluene-d8	7.894	98	379162	4.783	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.600%
43) trans-1,3-Dichloroprop...	8.177	79	50657	5.204	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	104.000%
46) 2-Hexanone-d5	8.627	63	183006	57.357	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	114.720%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	104245	5.735	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	114.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	141391	5.812	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	116.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110923\
Data File : VU056147.D
Acq On : 09 Nov 2023 18:03
Operator : MD/SY
Sample : 05328-07
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
A49W7

Quant Time: Nov 09 21:37:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 09 21:34:15 2023
Response via : Initial Calibration

