

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110923\
 Data File : VU056148.D
 Acq On : 09 Nov 2023 18:27
 Operator : MD/SY
 Sample : 05342-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AH3

Quant Time: Nov 09 21:38:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 09 21:34:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	377299	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	378458	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	161816	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	110898	4.735	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.800%
7) Chloroethane-d5	1.907	69	88602	5.575	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	111.400%
11) 1,1-Dichloroethene-d2	2.563	65	49430	4.791	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	95.800%
20) 2-Butanone-d5	4.618	46	258041	67.163	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	134.320%#
24) Chloroform-d	5.055	84	236182	5.298	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.000%
26) 1,2-Dichloroethane-d4	5.698	65	111569	5.684	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.600%
32) Benzene-d6	5.721	84	467136	5.139	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.800%
36) 1,2-Dichloropropane-d6	6.685	67	144535	5.402	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.000%
41) Toluene-d8	7.894	98	384804	4.620	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.400%
43) trans-1,3-Dichloroprop...	8.177	79	50299	4.918	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.400%
46) 2-Hexanone-d5	8.631	63	198598	59.244	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	118.480%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	107160	5.611	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	112.200%
66) 1,2-Dichlorobenzene-d4	12.190	152	139490	5.590	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.800%
Target Compounds						Qvalue
16) Methylene chloride	3.042	84	18800	0.676	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU110923\
Data File : VU056148.D
Acq On : 09 Nov 2023 18:27
Operator : MD/SY
Sample : 05342-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AH3

Quant Time: Nov 09 21:38:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR110723WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 09 21:34:15 2023
Response via : Initial Calibration

