

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU111020\
 Data File : VU041195.D
 Acq On : 11 Nov 2020 10:43
 Operator : SY/MD
 Sample : VU1110WBL02
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK21

Quant Time: Nov 11 13:43:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Nov 11 13:19:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	52291	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	55287	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	25057	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	20686	4.71	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.20%
7) Chloroethane-d5	1.92	69	19223	5.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.80%
11) 1,1-Dichloroethene-d2	2.58	63	32697	3.79	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.80%
20) 2-Butanone-d5	4.65	46	76407	52.82	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.64%
24) Chloroform-d	5.08	84	47319	6.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	122.00%
26) 1,2-Dichloroethane-d4	5.72	65	27069	5.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	112.40%
32) Benzene-d6	5.74	84	72822	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.60%
36) 1,2-Dichloropropane-d6	6.70	67	24150	5.02	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.40%
41) Toluene-d8	7.90	98	66437	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.20%
43) trans-1,3-Dichloropropene-	8.19	79	10247	4.91	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.20%
46) 2-Hexanone-d5	8.64	63	48998	43.61	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.22%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	24682	5.52	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	110.40%
64) 1,2-Dichlorobenzene-d4	12.19	152	24442	5.56	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU111020\
Data File : VU041195.D
Acq On : 11 Nov 2020 10:43
Operator : SY/MD
Sample : VU1110WBL02
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK21

Quant Time: Nov 11 13:43:14 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Nov 11 13:19:05 2020
Response via : Initial Calibration

