

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU111020\
 Data File : VU041198.D
 Acq On : 11 Nov 2020 13:10
 Operator : SY/MD
 Sample : L4646-10
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Nov 11 14:16:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Nov 11 13:19:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	53466	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	63634	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	23158	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	23719	5.28	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	105.60%
7) Chloroethane-d5	1.92	69	19617	5.68	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.60%
11) 1,1-Dichloroethene-d2	2.58	63	34506	3.92	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.40%
20) 2-Butanone-d5	4.65	46	80652	54.53	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.06%
24) Chloroform-d	5.08	84	44514	5.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.20%
26) 1,2-Dichloroethane-d4	5.72	65	27218	5.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.60%
32) Benzene-d6	5.74	84	77691	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.40%
36) 1,2-Dichloropropane-d6	6.70	67	23206	4.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.80%
41) Toluene-d8	7.90	98	59449	3.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.20%
43) trans-1,3-Dichloropropene-	8.18	79	9730	4.05	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.00%
46) 2-Hexanone-d5	8.64	63	49119	37.98	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	75.96%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	21158	4.11	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.20%
64) 1,2-Dichlorobenzene-d4	12.19	152	23286	5.73	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU111020\
Data File : VU041198.D
Acq On : 11 Nov 2020 13:10
Operator : SY/MD
Sample : L4646-10
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Nov 11 14:16:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Nov 11 13:19:05 2020
Response via : Initial Calibration

