


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU111020\
 Data File : VU041166.D
 Acq On : 10 Nov 2020 09:20
 Operator : SY/MD
 Sample : BFB09
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB09

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR102620WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Wed Nov 11 10:55:57 2020



AutoFind: Scans 2889, 2890, 2891; Background Corrected with Scan 2875

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.6	20506	PASS
75	95	30	80	56.3	53368	PASS
95	95	100	100	100.0	94770	PASS
96	95	5	9	6.3	6001	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	75.3	71389	PASS
175	174	5	9	7.7	5477	PASS
176	174	95	101	95.5	68154	PASS
177	176	5	9	6.2	4213	PASS

m/z	Abundance
36.00	1089.0
37.00	5884.0
38.10	4932.0
39.10	2175.0
44.00	954.0
45.00	1085.0
46.00	179.0
47.10	1110.0
48.00	681.0
49.10	4597.0
50.10	20936.0
51.10	6326.0
52.00	353.0
55.10	379.0
56.10	1717.0
57.10	3134.0
57.90	195.0
60.10	1038.0
61.10	5627.0
62.10	5804.0
63.10	4416.0
64.10	684.0
65.10	804.0
67.00	255.0
68.10	11057.0
69.10	10827.0
70.10	750.0
72.00	715.0
73.10	4823.0
74.10	18328.0
75.10	53224.0
76.10	4623.0
77.10	634.0
78.10	362.0
79.00	4860.0
80.00	1335.0
81.00	4895.0
82.00	1069.0
82.90	101.0
87.10	3278.0
88.00	3595.0
91.00	492.0
92.10	3135.0
93.10	4543.0
94.10	12386.0
95.10	93512.0
96.10	6024.0
97.10	149.0
104.00	612.0
104.90	163.0
106.00	522.0
106.80	110.0
113.10	120.0
114.90	114.0
116.00	339.0
117.00	943.0
117.90	427.0
119.00	788.0
128.00	487.0
129.00	221.0
130.00	295.0
131.00	224.0
135.00	214.0
136.90	188.0
141.00	1304.0
142.10	148.0
143.00	1747.0
144.10	114.0
145.00	441.0
146.00	149.0
147.90	246.0
150.00	151.0
155.00	176.0
157.00	177.0
160.90	160.0
171.10	233.0
172.10	1085.0
174.10	69704.0
175.10	5382.0
176.10	65752.0
177.10	4085.0
178.00	114.0

Instrument :
MSVOA_U
ClientSampleId :
BFB09

m/z	Abundance
36.00	1117.0
37.00	6024.0
38.10	4963.0
39.10	2239.0
39.80	107.0
44.00	954.0
45.00	1095.0
46.10	141.0
47.10	1156.0
48.00	689.0
49.10	4703.0
50.10	21248.0
51.10	6502.0
52.00	326.0
52.90	139.0
55.10	387.0
56.10	1699.0
57.00	3337.0
57.90	223.0
60.00	1114.0
61.10	5780.0
62.10	5998.0
63.10	4644.0
64.10	630.0
65.10	771.0
67.00	238.0
68.10	11420.0
69.10	11142.0
70.10	761.0
72.00	732.0
73.10	5183.0
74.10	18760.0
75.10	55256.0
76.10	4503.0
77.10	589.0
78.10	384.0
79.00	4916.0
80.00	1403.0
81.00	4850.0
82.00	1095.0
82.90	134.0
87.10	3467.0
88.00	3550.0
91.00	547.0
92.10	3124.0
93.10	4819.0
94.10	13090.0
95.10	98264.0
96.10	6207.0
97.10	181.0
104.00	575.0
104.90	143.0
106.00	545.0
106.90	110.0
113.00	136.0
114.90	137.0
116.00	370.0
117.00	984.0
117.90	474.0
119.00	746.0
128.00	496.0
129.00	184.0
130.00	347.0
131.00	177.0
135.00	215.0
136.90	214.0
141.00	1500.0
142.10	159.0
143.00	1769.0
144.10	120.0
145.00	405.0
145.90	163.0
147.00	114.0
147.90	287.0
150.00	131.0
155.00	204.0
156.90	179.0
160.90	185.0
171.10	247.0
172.10	1065.0
174.10	74016.0
175.10	5687.0
176.10	70488.0
177.10	4382.0
178.00	120.0

Instrument :
MSVOA_U
ClientSampleId :
BFB09

m/z	Abundance
36.00	1068.0
37.10	5484.0
38.10	4531.0
39.00	2068.0
40.00	108.0
44.00	877.0
45.00	930.0
46.20	119.0
47.10	1001.0
48.10	695.0
49.10	4380.0
50.10	19336.0
51.10	6115.0
52.10	245.0
52.90	165.0
55.10	342.0
56.10	1579.0
57.00	3181.0
57.90	202.0
60.00	1092.0
61.10	5305.0
62.10	5444.0
63.10	4463.0
64.10	521.0
65.10	717.0
67.00	216.0
68.10	10583.0
69.10	10323.0
70.00	743.0
72.00	637.0
73.10	4853.0
74.10	17352.0
75.10	51624.0
76.10	4015.0
77.10	528.0
78.00	368.0
79.00	4487.0
80.00	1325.0
81.00	4408.0
82.00	1029.0
82.90	132.0
87.10	3277.0
88.00	3071.0
91.00	528.0
92.10	2908.0
93.10	4748.0
94.10	12494.0
95.10	92536.0
96.10	5772.0
97.10	202.0
104.00	491.0
105.00	128.0
106.00	527.0
106.90	101.0
111.90	100.0
113.00	118.0
115.00	147.0
116.00	393.0
117.00	946.0
118.00	429.0
119.00	660.0
128.00	440.0
129.00	141.0
130.00	361.0
131.00	120.0
134.90	194.0
136.90	212.0
141.00	1460.0
142.10	149.0
143.10	1551.0
144.10	100.0
145.00	344.0
145.90	138.0
147.00	133.0
148.00	295.0
155.00	204.0
156.90	153.0
160.90	187.0
171.20	210.0
172.10	894.0
174.10	70448.0
175.10	5363.0
176.10	68224.0
177.10	4172.0
178.00	108.0

Instrument :
MSVOA_U
ClientSampleId :
BFB09

m/z	Abundance
44.00	104.0

Instrument :
MSVOA_U
ClientSampleId :
BFB09