

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
 Data File : VU045684.D
 Acq On : 10 Nov 2021 10:40
 Operator : SY/MD
 Sample : VU1110MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_U/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1110MBL01

Quant Time: Nov 10 15:32:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	128901	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	243701	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	237648	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	112216	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	104352	54.056	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.120%	
35) Dibromofluoromethane	5.292	113	82411	50.650	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.300%	
50) Toluene-d8	7.903	98	314745	52.022	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.040%	
62) 4-Bromofluorobenzene	10.636	95	114963	47.386	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.780%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
Data File : VU045684.D
Acq On : 10 Nov 2021 10:40
Operator : SY/MD
Sample : VU1110MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_U/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1110MBL01

Quant Time: Nov 10 15:32:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

