

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
 Data File : VU045685.D
 Acq On : 10 Nov 2021 14:08
 Operator : SY/MD
 Sample : VU1110WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1110WBL01

Quant Time: Nov 10 15:32:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	123597	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	231506	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	234345	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	113052	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.707	65	102620	55.440	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 110.880%	
35) Dibromofluoromethane	5.292	113	79296	51.303	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.600%	
50) Toluene-d8	7.903	98	301360	52.434	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.860%	
62) 4-Bromofluorobenzene	10.636	95	113029	49.043	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.080%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
Data File : VU045685.D
Acq On : 10 Nov 2021 14:08
Operator : SY/MD
Sample : VU1110WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1110WBL01

Quant Time: Nov 10 15:32:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

