

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
 Data File : VU045688.D
 Acq On : 10 Nov 2021 15:22
 Operator : SY/MD
 Sample : M4497-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RO-COMP-1

Quant Time: Nov 10 16:37:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	124873	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	235560	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	240629	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	115277	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	101564	54.309	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.620%	
35) Dibromofluoromethane	5.292	113	78972	50.214	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.420%	
50) Toluene-d8	7.902	98	312944	53.512	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.020%	
62) 4-Bromofluorobenzene	10.635	95	115473	49.241	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.480%	
Target Compounds						Qvalue
16) Acetone	2.626	43	16266	17.086	ug/l	100
20) Methylene Chloride	3.051	84	6554	3.962	ug/l	94
43) Isopropyl Acetate	5.915	43	18648	4.621	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
Data File : VU045688.D
Acq On : 10 Nov 2021 15:22
Operator : SY/MD
Sample : M4497-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RO-COMP-1

Quant Time: Nov 10 16:37:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

