

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
 Data File : VU045689.D
 Acq On : 10 Nov 2021 15:45
 Operator : SY/MD
 Sample : M4497-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VNJ-203

Quant Time: Nov 10 16:38:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	125953	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	243436	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	248112	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	118823	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	103656	54.952	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.900%	
35) Dibromofluoromethane	5.292	113	80858	49.750	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.500%	
50) Toluene-d8	7.899	98	321582	53.210	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.420%	
62) 4-Bromofluorobenzene	10.636	95	120902	49.888	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.780%	
Target Compounds						Qvalue
16) Acetone	2.626	43	14204	14.793	ug/l	100
20) Methylene Chloride	3.051	84	6805	4.079	ug/l	97
43) Isopropyl Acetate	5.915	43	18077	4.334	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
Data File : VU045689.D
Acq On : 10 Nov 2021 15:45
Operator : SY/MD
Sample : M4497-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VNJ-203

Quant Time: Nov 10 16:38:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

