

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
 Data File : VU045697.D
 Acq On : 10 Nov 2021 18:49
 Operator : SY/MD
 Sample : M4593-02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 WASTE-108

Quant Time: Nov 11 02:31:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	126824	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	246994	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	252482	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	123963	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	105359	55.472	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.940%
35) Dibromofluoromethane	5.292	113	81713	49.552	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.100%
50) Toluene-d8	7.899	98	323412	52.742	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.480%
62) 4-Bromofluorobenzene	10.636	95	122175	49.687	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.380%
Target Compounds						Qvalue
16) Acetone	2.626	43	18898	19.546	ug/l	96
20) Methylene Chloride	3.051	84	8418	5.011	ug/l	96
43) Isopropyl Acetate	5.912	43	17723	4.188	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
Data File : VU045697.D
Acq On : 10 Nov 2021 18:49
Operator : SY/MD
Sample : M4593-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
WASTE-108

Quant Time: Nov 11 02:31:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

