

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
 Data File : VU045704.D
 Acq On : 10 Nov 2021 21:54
 Operator : SY/MD
 Sample : VU1110WBL02
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1110WBL02

Quant Time: Nov 11 02:33:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	122014	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	222902	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	223983	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	112262	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	97415	53.311	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.620%	
35) Dibromofluoromethane	5.292	113	75697	50.865	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.720%	
50) Toluene-d8	7.902	98	291073	52.599	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.200%	
62) 4-Bromofluorobenzene	10.635	95	110645	49.861	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.720%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
Data File : VU045704.D
Acq On : 10 Nov 2021 21:54
Operator : SY/MD
Sample : VU1110WBL02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1110WBL02

Quant Time: Nov 11 02:33:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

