

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
 Data File : VU045709.D
 Acq On : 11 Nov 2021 00:13
 Operator : SY/MD
 Sample : M4594-11DL 20X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 91BR-20211108DL

Quant Time: Nov 22 05:02:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	126290	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	232020	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	241557	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	120361	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	101281	53.550	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.100%	
35) Dibromofluoromethane	5.292	113	79451	51.289	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.580%	
50) Toluene-d8	7.899	98	307120	53.317	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.640%	
62) 4-Bromofluorobenzene	10.636	95	118024	51.097	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.200%	
Target Compounds						Qvalue
4) Vinyl Chloride	1.604	62	18239	9.994	ug/l	100
27) cis-1,2-Dichloroethene	4.671	96	58708	34.580	ug/l	90
44) Trichloroethene	6.546	130	4797	2.992	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
Data File : VU045709.D
Acq On : 11 Nov 2021 00:13
Operator : SY/MD
Sample : M4594-11DL 20X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
91BR-20211108DL

Quant Time: Nov 22 05:02:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

