

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
 Data File : VU045710.D
 Acq On : 11 Nov 2021 00:36
 Operator : SY/MD
 Sample : M4594-02DL 40X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 15BR-20211108DL

Quant Time: Nov 11 02:35:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.375	168	127108	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	234147	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	246016	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.816	152	122654	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	103656	54.453	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.900%
35) Dibromofluoromethane	5.292	113	80116	51.249	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.500%
50) Toluene-d8	7.903	98	309820	53.298	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.600%
62) 4-Bromofluorobenzene	10.636	95	119448	51.243	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.480%
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.607	62	6265	3.411	ug/l	96
27) cis-1,2-Dichloroethene	4.671	96	93384	54.651	ug/l	90
44) Trichloroethene	6.546	130	109116	67.439	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
Data File : VU045710.D
Acq On : 11 Nov 2021 00:36
Operator : SY/MD
Sample : M4594-02DL 40X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
15BR-20211108DL

Quant Time: Nov 11 02:35:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

