

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
 Data File : VU045717.D
 Acq On : 11 Nov 2021 03:18
 Operator : SY/MD
 Sample : M4594-03DL 50X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 20BR-20211108DL

Quant Time: Nov 11 13:08:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	122034	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	220209	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	226370	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.815	152	112113	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	96721	52.922	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.840%	
35) Dibromofluoromethane	5.292	113	75152	51.116	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.240%	
50) Toluene-d8	7.899	98	292484	53.500	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.000%	
62) 4-Bromofluorobenzene	10.635	95	109312	49.863	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.720%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.604	62	23531	13.343	ug/l	100
27) cis-1,2-Dichloroethene	4.671	96	126652	77.202	ug/l	90
44) Trichloroethene	6.546	130	31449	20.667	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
Data File : VU045717.D
Acq On : 11 Nov 2021 03:18
Operator : SY/MD
Sample : M4594-03DL 50X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
20BR-20211108DL

Quant Time: Nov 11 13:08:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

