

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
 Data File : VU045720.D
 Acq On : 11 Nov 2021 04:27
 Operator : SY/MD
 Sample : M4594-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 22BR-20211108

Quant Time: Nov 11 13:09:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	119025	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	221561	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	231185	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	114554	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	97235	54.549	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.100%	
35) Dibromofluoromethane	5.292	113	76731	51.872	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.740%	
50) Toluene-d8	7.899	98	293254	53.313	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.620%	
62) 4-Bromofluorobenzene	10.635	95	111477	50.540	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.080%	
Target Compounds						
					Qvalue	
16) Acetone	2.629	43	1396	1.538	ug/l	97
27) cis-1,2-Dichloroethene	4.671	96	21955	13.721	ug/l	88
44) Trichloroethene	6.546	130	13001	8.492	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
Data File : VU045720.D
Acq On : 11 Nov 2021 04:27
Operator : SY/MD
Sample : M4594-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
22BR-20211108

Quant Time: Nov 11 13:09:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

