

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
 Data File : VU045723.D
 Acq On : 11 Nov 2021 05:36
 Operator : SY/MD
 Sample : M4594-15
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FB01-20211108

Quant Time: Nov 11 13:10:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.378	168	118081	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	218280	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	227402	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	112803	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.706	65	96744	54.707	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.420%	
35) Dibromofluoromethane	5.292	113	74605	51.193	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.380%	
50) Toluene-d8	7.899	98	287061	52.972	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.940%	
62) 4-Bromofluorobenzene	10.635	95	109379	50.335	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.660%	
Target Compounds						
16) Acetone	2.626	43	3884	4.315	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
Data File : VU045723.D
Acq On : 11 Nov 2021 05:36
Operator : SY/MD
Sample : M4594-15
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FB01-20211108

Quant Time: Nov 11 13:10:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

