

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
 Data File : VU045725.D
 Acq On : 11 Nov 2021 06:22
 Operator : SY/MD
 Sample : M4600-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FM786-N34145

Quant Time: Nov 11 13:11:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 05 13:35:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.379	168	118675	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.253	114	222220	50.000	ug/l	0.00
63) Chlorobenzene-d5	9.420	117	231929	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.812	152	115631	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.703	65	97831	55.045	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 110.080%	
35) Dibromofluoromethane	5.295	113	76196	51.357	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.720%	
50) Toluene-d8	7.899	98	294628	53.404	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 106.800%	
62) 4-Bromofluorobenzene	10.636	95	112663	50.927	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.860%	
Target Compounds						Qvalue
16) Acetone	2.629	43	4391	4.853	ug/l	99
30) Chloroform	5.092	83	4332	1.606	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
Data File : VU045725.D
Acq On : 11 Nov 2021 06:22
Operator : SY/MD
Sample : M4600-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FM786-N34145

Quant Time: Nov 11 13:11:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 05 13:35:16 2021
Response via : Initial Calibration

