



Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111021\
 Data File : VU045702.D
 Acq On : 10 Nov 2021 20:45
 Operator : SY/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U110521W.M
 Title : SW846 8260
 Last Update : Fri Nov 05 13:35:16 2021



AutoFind: Scans 2890, 2891, 2892; Background Corrected with Scan 2878

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.4	6072	PASS
75	95	30	60	54.1	17813	PASS
95	95	100	100	100.0	32944	PASS
96	95	5	9	6.6	2189	PASS
173	174	0.00	2	1.1	270	PASS
174	95	50	100	74.5	24541	PASS
175	174	5	9	7.4	1804	PASS
176	174	95	101	99.1	24328	PASS
177	176	5	9	6.0	1460	PASS