

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111022\
 Data File : VU051805.D
 Acq On : 10 Nov 2022 17:00
 Operator : JC/MD
 Sample : N5571-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBMW4

Quant Time: Nov 11 00:40:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 11 00:38:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	196515	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	206003	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	76552	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	68029	4.596	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.000%
7) Chloroethane-d5	1.909	69	67269	4.358	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.200%
11) 1,1-Dichloroethene-d2	2.565	65	28211	5.252	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	105.000%
20) 2-Butanone-d5	4.617	46	203349	37.937	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	75.880%
24) Chloroform-d	5.060	84	149165	4.307	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.200%
26) 1,2-Dichloroethane-d4	5.700	65	80194	4.286	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.800%
32) Benzene-d6	5.726	84	300142	4.617	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.400%
36) 1,2-Dichloropropane-d6	6.687	67	101398	4.461	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.200%
41) Toluene-d8	7.896	98	234574	4.588	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.800%
43) trans-1,3-Dichloroprop...	8.179	79	31282	4.501	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.000%
46) 2-Hexanone-d5	8.629	63	154733	36.981	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	73.960%
56) 1,1,2,2-Tetrachloroeth...	10.755	84	91273	4.382	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.600%
66) 1,2-Dichlorobenzene-d4	12.192	152	84015	5.980	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	119.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU111022\
Data File : VU051805.D
Acq On : 10 Nov 2022 17:00
Operator : JC/MD
Sample : N5571-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBMW4

Quant Time: Nov 11 00:40:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR102822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 11 00:38:34 2022
Response via : Initial Calibration

